

Test Results

Respirators help protect against certain particulate contaminants but do not eliminate exposure to or the risk of contracting any disease or infection. Misuse of a respirator can result in exposures that can result in sickness or death.

The test results below indicate the performance of the filtration material specific to the respirator tested. The materials were tested using specialized equipment and protocols, similar but not equivalent to the NIOSH standard procedure TEB-APR-STP-0059; reported filtration efficiencies may differ accordingly. These results do not guarantee that all respirators produced by a manufacturer will have a similar level of filtration efficiency or a similar level of resistance to fluid penetration. These results do not imply a fitness for a particular use or compliance with any federal or international standards. These respirators were not tested in an accredited National Institute for Occupational Safety and Health (NIOSH) testing laboratory, and thus are not NIOSH approved or certified. We have worked with the Advanced Functional Fabrics of America, Inc. (AFFOA) to obtain respirator performance characterization. AFFOA coordinated testing at MIT Department of Chemical Engineering and MIT Lincoln Laboratory where comparable conclusions about mask quality were obtained. The Filtration efficiency measurements listed below were taken by the Massachusetts Institute of Technology. The resistance to fluid penetration measurements below were taken by MIT Lincoln Laboratory.

According to [ASTM F 1862](#), surgical masks are tested on a pass/fail basis at three velocities corresponding to the range of human blood pressure (80, 120, 160 mm Hg). Fluid resistance may be claimed if the device passes ASTM F1862 at the corresponding level. Surgical masks that show passing results at higher velocities are more fluid resistant. The test results below indicate the performance of the specific mask tested.








POWECOM

Test date – 4/18/20 - Filtration efficiency: 90.6%



Test date – 4/17/20 – Filtration efficiency: 92 %


[Visit CDC website for additional information on this product](#)


	<p>Jiangsu Nanfang Medical CO. LTD Test date – 4/17/20 – Filtration efficiency: 90.3% Test Date – 4/24/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p>
	<p>Foshan Wei hui Labor Protection Products Co., Ltd Test date – 4/17/20 – Filtration efficiency: 84.2% Test Date – 4/24/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through, but partial layer penetration along seams</p>
	<p>Unlabeled – plastic bag with KN95 on the mask Test date – 4/17/20 – Filtration efficiency: 28.1%</p>
	<p>KN-95 Disposable Face Mask Test date – 4/16/20 – Filtration efficiency: 52.3% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through, but partial layer penetration along seams</p>
	<p>SHUYIAN Test date – 4/16/20 – Filtration efficiency: 87.6%</p>

 <p>Protective mask PinzTec 10 PACK #2</p>	<p>PinzTec Test date – 4/16/20 – Filtration efficiency: 83.6%</p>
 <p>SDI 丽洁 自吸式防颗粒物呼吸器 KN95 GB2626-2006</p>	<p>SDI Folding Anti-Particulate respirator Test date – 4/16/20 – Filtration efficiency: 91.4%</p>
 <p>MEDICAL FACE MASK Cover Cover EO 974</p>	<p>ESOUND Med Test date – 4/23/2020 – Filtration efficiency: 88.2% Test date – 4/27/2020 – Filtration efficiency: 86% Test Date – 4/29/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p>
 <p>Donation Dräger X-plore 1800 V FFP2 10x</p>	<p>Dräger Test date – 4/24/2020 – Filtration efficiency: 95% Test Date – 4/29/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p>
 <p>KN95 Respirator 50</p>	<p>AOXING Test date – 4/24/2020 – Filtration efficiency: 87.8% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through, but partial layer penetration along seams</p>

	<p>Dasheng Model DTC3X Test Date – 4/24/2020 – Filtration efficiency: 88.7% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p> <p>Visit CDC website for additional information on this product</p>
	<p>Americares Emergency Response BYD CARE Test Date – 4/24/2020 – Filtration efficiency: 89% Test Date – 4/29/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p>
	<p>KN95 (FFP2) Test Date – 4/24/2020 – Filtration efficiency: 70% Test Date – 4/29/2020 – Penetration Resistance Test: Failed</p>
	<p>GONGRENCHENGPING Respiratory Face Mask Test Date – 4/24/2020 – Filtration efficiency: 85.9% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); but partial layer penetration along stitches</p>
	<p>LVJIANGNAN Test Date – 4/24/2020 – Filtration efficiency: 91.4% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p>

	<p>DASHENG MODEL DTC3W Test Date – 4/26/2020 – Filtration efficiency: 91.8% Test Date – 4/27/2020 – Penetration Resistance Test: Failed</p>
	<p>Saiertech Test Date – 4/26/2020 – Filtration efficiency: 85.9% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through, but partial layer penetration along seams</p>
	<p>OANY Test Date – 4/26/2020 – Filtration Efficiency: 62.3% Test Date – 4/29/2020 – Penetration Resistance Test: Failed</p>
	<p>Gerber Outerware Test Date – 4/26/2020 – Filtration efficiency: 77.3% Test Date – 4/29/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); no soak through</p>
	<p>PURVIGOR Test Date – 4/27/2020 – Filtration efficiency: 73.4% Test Date – 4/29/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); but partial layer penetration along stitches Visit CDC website for additional information on this product</p>

	<p>JCH Industrial, Hong Kong Test Date – 4/27/2020 – Filtration efficiency: 86.7% Test Date – 4/29/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); but partial layer penetration along stitches</p>
	<p>Lamdown Test Date – 4/27/2020 – Filtration efficiency: 87.4% Test Date – 4/27/2020 – Penetration Resistance Test: Passed High Standard (Level 3, 160mmHg); but partial layer penetration along stitches</p>
	<p>N/A Test Date – 4/27/2020 – Filtration efficiency: 62.6% Test Date – 4/29/2020 – Penetration Resistance Test: Failed</p>
	<p>Yiwu Hongbiao Garments Co. Ltd Test Date – 4/29/2020 – Filtration Efficiency: 91.1%</p>
	<p>3M KN-95 Campeco, Inc. Test Date – 4/29/2020 – Filtration Efficiency: 95.6%</p>



Makrite
 Test Date – 4/26/20 – Filtration Efficiency: 91.6%



Safe Secure
 Test Date – 4/29/2020 - Filtration Efficiency: 89.8%



GB2626-2006 KN95
 Test Date – 4/29/2020 - Filtration Efficiency: 78.9%



PM 2.5, 5-pack
 Test Date – 4/29/2020 - Filtration Efficiency: 73.20%



HG, 5-pack
 Test Date – 4/29/2020 - Filtration Efficiency: 51.10%



ONE MIX
 Test Date – 4/29/2020 - Filtration Efficiency: 85.3%



Huran
 Test Date – 4/29/2020 - Filtration Efficiency: 93.00%








Ace Neale
Test Date – 4/29/2020 - Filtration Efficiency: 92.0%



Guan Hua Item 9801
Test Date – 4/29/2020 - Filtration Efficiency: 84.7%



Linxia
Test Date – 4/29/2020 - Filtration Efficiency: 89.2%

	<p>KANGERDA Test Date – 4/29/2020 – Filtration Efficiency: 70.5%</p>
	<p>DROMEX (Campo Inc.) Test Date – 4/30/2020 – Filtration Efficiency: 80.9%</p>
	<p>S. Sangso Test Date – 4/30/2020 – Filtration Efficiency: 90.0%</p>
	<p>ZHIYI Test Date – 5/5/2020 – Filtration Efficiency: 82.4 %</p>
	<p>Chun Lam Group Test Date – 5/5/2020 – Filtration Efficiency: 86.0%</p>



DR. MFYAN
 Test Date – 5/5/2020 – Filtration Efficiency: 45.5% %



ALLMAN
 Test Date – 5/5/2020 – Filtration Efficiency: 92.4 %



Doctor Hua (Jiayuan Technology)
 Test Date – 5/5/2020 – Filtration Efficiency: 63.4 %



Shenzhen Zhiyou
 Test Date – 5/5/2020 – Filtration Efficiency: 61.8%



Shanghai Huaxiang Woolen Dressing
 Huaxiang Woolen
 Test Date – 5/5/2020 – Filtration Efficiency: 84.1 %

	<p>China Southern Test Date – 5/5/2020 – Filtration Efficiency: 72.8 %</p>
	<p>Ocean State Donation Test Date – 5/5/2020 – Filtration Efficiency: 86.0 %</p>
	<p>FFP2 NK International (King Year Printing & Packaging Co. Ltd.) Test Date – 5/5/2020 – Filtration Efficiency: 75.4%</p>
	<p>Model:YH006 (Joint China Ltd.) Test Date – 5/5/2020 – Filtration Efficiency: 81.6%</p>
	<p>BIWEIKANG Test Date: 5/8/2020 – Filtration Efficiency: 85.5%</p>



Easewell (Easepal Ind. Co. Ltd.)
 Test Date: 5/8/2020 – Filtration Efficiency: 81.7%



Makrite (SEKURA)
 Test Date: 5/8/2020 – Filtration Efficiency: 92.6%



EATA
 Test Date: 5/8/2020 – Filtration Efficiency: 73.2%



Uniair (SAN HUEI United Co. Ltd)
 Test Date: 5/8/2020 – Filtration Efficiency: 88.9%



Xier
 Test Date: 5/12/2020 – Filtration Efficiency: 62.0%



Spectrum 50-pack
 Test Date: 5/12/2020 – Filtration Efficiency: 72.5%




Eko
 Test Date: 5/12/2020 – Filtration Efficiency: 41.9%



San Hwei United Co., Ltd. SH9550
 Test Date: 5/12/2020 – Filtration Efficiency: 86.4%



In Jeolly
 Test Date: 5/12/2020 – Filtration Efficiency: 83.4%



Fujian Meinkind Baby Products, Ltd
 Test Date: 5/12/2020 – Filtration Efficiency: 71.1%



50 pack donation
Test Date: 5/12/2020 – Filtration Efficiency: 91.1%



Aumacom
Test Date: 5/12/2020 – Filtration Efficiency: 48.5%



XAM-Med
Test Date: 5/12/2020 – Filtration Efficiency: 85.7%



Photographs of Penetration Resistance Tests



Pass



Pass
with partial penetration along seams



Fail



Presentation Name: 8
Author Initials: MIDDYY

LINCOLN LABORATORY
MASSACHUSETTS INSTITUTE OF TECHNOLOGY

This document will be updated as additional masks and samples are tested